

Research MRI Safety Committee Standard

REPORTING REQUIREMENTS

Category: Magnetic Resonance Imaging (MRI) Safety

Procedure #: MR.SOP.10

Applies to: Investigators, study personnel, Medical College of Wisconsin (MCW) staff

PURPOSE:

The mission of the Research MRI Safety Committee at the Medical College of

Research MRI Safety Committee and administration

2. Timeframe of the report:
 - a. Within 1 hour of occurrence, if the damage/failure affects the primary function of the MRI System
 - b. Within 8 hours of occurrence, if the damage/failure affects the ancillary function of the MRI System
 - c. Within 24 hours of occurrence, if the damage/failure affects components of the MRI System that do not affect the system's ability to image
3. Must be reported by: Study Personnel if discovered during a study

Approval Date: 02/15/2024
Effective Date: 02/15/2024
Review Date: 02/15/2024
Revision History: 01/2009, 12/17/2012, 11/13/2014